

Simultaneous and sequential X-ray ptychography and fluorescence microscopy at the Australian Synchrotron

M. W. M. Jones^{a,b}, K. Elgass^c, N. Phillips^{b,d,e}, D. J. Vine^f, G. A. van Riessen^d, M. D. de Jonge^a

^aAustralian Synchrotron, Clayton, 3168, Australia, ^bARC Centre of Excellence for Advanced Molecular Imaging, ^cMonash Institute of Medical Research, Monash Micro Imaging, Monash University, Clayton, 3186, Australia, ^dDepartment of Physics, La Trobe University, Bundoora, 3086, Australia, ^eCSIRO Manufacturing Flagship, Parkville, 3052, Australia, ^fX-ray Science Division, Advanced Photon Source, Argonne National Laboratory, Argonne, 60439, USA

Abstract

X-ray ptychography have seen many recent advances including sample delivery [1], data collection and analysis [2, 3], and data interpretation [4, 5]. Some of these advances have allowed water-window ptychography to be applied to hydrated specimens [1, 6] allow biological specimens to be imaged with high contrast in close to their natural environment. Other advances, such as on-the-fly scanning and rapid data processing offer improvements in the efficiency of ptychography that make simultaneous application of other scanning microscopy methods such as X-ray fluorescence microscopy experimentally viable [7]. For some specimens, the experimental requirements of these two methods overlap, and the measurements are best taken simultaneously [8]. However, for many cellular applications, the ideal conditions for each method are very different, requiring a sequential set of optimized measurements for ideal results [9]. In this presentation, we outline the case for each type of measurement and discuss recent applications of these two methods at the X-ray Fluorescence Microscopy and Soft X-ray Imaging beamlines at the Australian Synchrotron.

References

- [1] D. Nam, et al., Physical Review Letters 110 (2013) 098103.
- [2] P. M. Pelz, et al., Applied Physics Letters 105 (2014) 251101.
- [3] J. Deng, et al., Optics Express 23 (2015) 5438.
- [4] D. A. Shapiro, et al., Nature Photonics 8 (2014) 765.
- [5] M. W. M. Jones, et al., Scientific Reports 4 (2014) 6796.
- [6] M. W. M. Jones, et al., Optics Express 21 (2013) 32151.
- [7] J. Deng, et al., Proceedings of the National Academy of Science 112 (2015) 2314.
- [8] D. J. Vine, et al., Optics Express 20 (2012) 18287.
- [9] M. W. M. Jones, et al., in preparation